
Temporal Imaging System Demonstrates 103x Magnification and 300 fs Resolution

C. V. Bennett B. H. Kolner

This paper was prepared for submittal to the Conference on Lasers and Electro-Optics Baltimore, MD May 23-28, 1999

November 11, 1998



DISCLAIMER

This document was prepared as an account of work sponsored by an agency of the United States Government. Neither the United States Government nor the University of California nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States Government or the University of California. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States Government or the University of California, and shall not be used for advertising or product endorsement purposes.

Temporal Imaging System Demonstrates 103x Magnification and 300 fs Resolution

Corey V. Bennett

Dept. of Electrical Engineering, University of California, Los Angeles
and

Lawrence Livermore National Laboratory, P.O. Box 808, L-174, Livermore, California, 94550

Phone: (925) 422-9394, E-mail: cvbennett@llnl.gov

Brian H. Kolner

Department of Applied Science, 228 Walker Hall, University of California, Davis, Davis, California, 95616

Phone: (530) 754-4370, E-mail: kolner@leorg.ucdavis.edu

Abstract

A temporal imaging system obtaining 103x magnification and 300 fs resolution was demonstrated. This system consists of dispersive delay lines and a time lens produced by sum-frequency generation with a chirped pump.

Temporal Imaging System Demonstrates 103x Magnification and 300 fs Resolution

Corey V. Bennett

Dept. of Electrical Engineering, University of California, Los Angeles
and

Lawrence Livermore National Laboratory, P.O. Box 808, L-174, Livermore, California, 94550

E-mail: cvbennett@llnl.gov

Brian H. Kolner

Department of Applied Science, 228 Walker Hall, University of California, Davis,

Davis, California, 95616

E-mail: kolner@leorg.ucdavis.edu

The study of transient phenomena with ultrafast detail necessitates the development of new measurement techniques. Temporal imaging^{1,2} is an ultrafast pulse manipulation technique that can expand an input waveform while maintaining its original profile thereby allowing data to be recorded with conventional technology. Based on an analogy that exists between paraxial diffraction and narrowband dispersion, the system consists of dispersive delay lines on the input and output combined with a quadratic temporal phase modulation (i.e. a linear frequency chirp) between them acting as a time lens, as shown in Fig. 1. We have demonstrated temporal imaging with 103x magnification and 300 fs resolution.

The f-number of a time lens and thus the system resolution depends inversely on the bandwidth imparted by the modulation process; $f^{\#} = \omega_0/\Delta\omega$. The required bandwidth was obtained from an 87 fs (5.0 THz) pulse from a Kerr-lens modelocked Ti:Sapphire laser which was sent through a grating pair dispersive delay line³ to obtain the desired linear chirp. When mixed in a 500 μ m thick BBO crystal with the dispersed input signal the chirped

C. V. Bennett, B. H. Kolner, "Temporal Imaging System Demonstrates 103x Magnification and 300 fs Resolution" pump pulse imparts its phase and amplitude profile to the up-converted signal, producing the necessary time lens action⁴ for an f-number of 82.

Given the pump pulse chirp rate is $d\omega/d\tau$ we may define a focal group delay dispersion (GDD) as $\phi_f'' = -(d\omega/d\tau)^{-1}$ which is the amount of GDD required to remove the phase imparted by the lens. When we cascade an input GDD (ϕ_1'') , the time lens, and an output GDD (ϕ_2'') , in the proper balance to satisfy the imaging condition, $1/\phi_1'' + 1/\phi_2'' = 1/\phi_f''$, a temporal image with magnification $M = -\phi_2''/\phi_1''$ is created.

Since GDD can be positive or negative it is possible to create temporal images with a single lens having either positive or negative magnification. We designed our system for M = +100 using $\phi_1'' = +0.17606$ ps² and $\phi_2'' = -17.606$ ps², achieved with diffraction grating dispersive delay lines.^{3,5} The pump chirp was also obtained with a grating pair and resulted in a focal dispersion of $\phi_f'' = +0.17784$ ps².

Higher order phase terms produce aberrations in temporal imaging systems. These occur in the spectral phase of the dispersion networks as well as the temporal phase of the time lens. We have discovered that the dominant result of higher order spectral phase in the input of the system is to distort the shape of the system's impulse response whereas when it is in the output it alters the arrival time of the impulse response. Time lens aberrations can produce a variety of effects, including aberrations similar to coma and spherical aberrations in spatial systems. Although our system was configured to minimize the effects of these aberrations, we plan a systematic study of their effects for the future.

A series of temporal images recorded with a 40 GHz photodiode and sampling oscilloscope using a two pulse input sequence are shown in Fig. 2. The delay between the input pulses was changed by 667 fs for each measurement. A change of 68.8 ps was measured in the output image, demonstrating a magnification of M = +103. With smaller delays two pulses can still be resolved with approximately 300 fs between them.

In conclusion, we have demonstrated a temporal imaging system with +103x magnification and 300 fs resolution. In addition, since the principles of temporal imaging do not rely on sampling, it should be suitable for extending the range of single-shot waveform recording. It is expected that a scaling of this technology will lead to the realization of a new class of long record length, single transient recorders with ultrafast resolution.

This research was supported by the Lawrence Livermore National Laboratory Photonics Group under LDRD grant No. 98-ERD-027, the National Science Foundation, the ATRI program of the US Air Force, and the David and Lucile Packard Foundation.

References

- 1. B. H. Kolner and M. Nazarathy, Opt. Lett., **14**, 630 (1989), and erratum **15**, 655 (1990).
- 2. B. H. Kolner, IEEE J. Quantum Electron., **30**, 1951 (1994).
- 3. E. B. Treacy, IEEE J. Quantum Electron., **QE-5**, 454 (1969).
- 4. C. V. Bennett, R. P. Scott, B. H. Kolner, Appl. Phys. Lett., **65**, 2513 (1994).
- 5. O. E. Martinez, J. P. Gordon, and R. L. Fork, J. Opt. Soc. Am. A, 1, 1003 (1984).

Figures

- Fig. 1. An up-conversion temporal imaging system with positive magnification. In temporal imaging narrowband dispersion plays the role of paraxial diffraction and imparting a quadratic temporal phase (or linear frequency chirp) through sum-frequency generation acts as a time lens. The input and output GDD are opposite in sign, similar to "virtual" spatial imaging systems, but in this case they produce a real measurable image.
- Fig. 2. A series of temporal images recorded at the output of the system with changing delay between the two input pulses, $\Delta \tau_{\rm in}$, in Fig. 1. The input waveform is made up of two 87 fs pulses. Pulse #2 is delayed by an additional 667 fs with respect to pulse #1 between each measurement. A fit to the position of pulse #2 gives a magnification of M = +103 with an error of 73 fs rms referred to the input. The ringing after each pulse is due to the impulse response of the photodiode, **not** aberrations in the temporal imaging system.

